

# Medical Examination Gloves

## 一、 Product name, model and specification

Product name : Medical examination gloves

Model : power/pitting surface、 power/smooth surface、 power-free/pitting surface、 power-free/smooth surface.

Specification : below 6 and 6 ( Special trumpet )、 6.5 ( S )、 7 and 7.5 ( )、 8 and 8.5 ( L )、 above 9 and 9

## 二、 Name, address, production address, contact information and after-sales service unit of the registered person/manufacturer

1、 Name: Hunan Zhenheyikang Medical Instrument Co.,LTD

2、 Address: No.6 Building Jingxiang Energy, No.55 Xianguang east road, Gaoxin district Xiangtan, Hunan

3、 Production address:NO.6 Building, NO.8 Building,5th floor Jingxiang Energy, No.55 Xianguang east road,Gaoxin district Xiangtan, Hunan

4、 Contact number : +86 17711627084

5、 After-sales service unit: Hunan Zhenheyikang Medical Instrument Co.,LTD

## 三、 Production Registration Certificate No: Xiangtan Medical Products Administration Production Registration NO 20140004

## 四、 Product Registration Certificate No、 Product Technical Requirements No:

XTFDA Certified 20210062

## 五、 Product description, intended use

The product is made of natural rubber latex, nitrile rubber latex, neoprene rubber latex, styrene butadiene rubber solution, styrene butadiene rubber rubber emulsion or thermoplastic elastic solution. Have sufficient strength and barrier performance. Non-sterile provided, single use.

Wearing it on doctor's hands or fingers when the patient's condition were examined or palpated

## 六、 Use method : Open the bag can be used, directly to treat the wound

## 七、 Contraindications, precautions, warnings and reminders

Product contraindications: None

Damaged packaging prohibited use, expired products prohibited use

Disposable use, please destroy after use

Avoid high temperature, humidity, direct sunlight

## 八、 Interpretation of graphics, symbols, abbreviations and other contents used in medical device labels



一次性使用



参考使用说明书



包装破损切勿使用

## 九、 Installation and use instructions or illustrations

This product does not need to be installed, open the bag can be used.

## 十、 Product maintenance and maintenance methods, special storage, transportation conditions, methods

Products should be stored in a non-corrosive gas and well-ventilated room, transported by ordinary transportation means, keep clean and pollution-free, rain and snow immersion.

## 十一、 Date of production, shelf life or expiration date

Valid for three years, production batch number, production date, expiry date please refer to the small package or qualification label.

## 十二、 Accessories list, including accessories, appendix, lossitage replacement cycle and replacement method description

This product has no accessories

## 十三、 Date and version of preparation or revision of the specification

This specification was revised in May 2023, version 2023.